

This equipment has been specifically designed for primary and secondary laboratories and has been supplied world-wide to researchers working in many fields of thermometric investigations and calibration.



DISPLAY SPECIFICATION

- Range : -30° C. to 50° C.
- Display : Dual 4 digit digital display
- Display resolution : 0.1° C.
- Temp. Setting : With push-button data entry
- Method of control make : PID controller
- Make : MIEPL

GENERAL SPECIFICATION

- System accuracy : Better than ± 0.3% of reading or ± 3° C.
- Stability : Better than ± 0.2° C.
- Set point achieve time : Approx. 60 Minutes
- Stabilization time : Approx. 30 minutes after reaching set value
- Media : Kerosene
- Operating Temp. : 0 - 20 ° C

ELECTRICAL SPECIFICATION

- Power consumption : 2000 Watts approx.
- Power supply : 230 Vac, 50 Hz
- Fuse rating : 15A, 250 Vac
- Cooling system : Air cooling through fans

PHYSICAL SPECIFICATION

- Dimensions : 13" (W) X 15" (H) X 37" (L)
- Weight : 30 Kg. approx.
- Well adapter M. O. C. : Aluminium
- Well Diameter : 90 mm
- Well Depth : 150 mm
- Well Volume : 600 ml
- Stirring : By mechanical stirrer of PMDC

MILLENNIUM INSTRUMENTS LIMITED

H.O. : 101, 'Dwarkashish', Main Vikas Marg, Shakarpur, Delhi-92 (INDIA)
Tel. : +91-11-2201-4325, 4327, 65094516. **Fax.:** 22014326
Website : www.mil.org.in , www.miepl.com, **E-mail :** miepl@miepl.com

Works : F-12, Site-B, UPSIDC Ind. Area Surajpur, Greater Noida, Gautam Budh Nagar-201306 (U.P.) **E-mail :** sales@miepl.com, info@miepl.com