

This equipment has been specifically designed for primary and secondary laboratories and has been supplied world-wide to researchers working in many fields of thermometric investigations and calibration.



DISPLAY SPECIFICATION

- Range : Ambient +5° C. to 400° C.
- Display : Dual 4 digit digital display
- Display resolution : 0.1° C.
- Temp. Setting : With push-button data entry
- Cooling Time : 60 Minutes from 400° C. to 100° C.
- Method of control make : PID controller
- Make : MIEPL

GENERAL SPECIFICATION

- System accuracy : Better than ± 0.3% of reading
- Stability : Better than ± 0.3° C.
- Set point achieve time : Approx. 45 Minutes
- Stabilization time : Approx. 30 minutes after reaching set value
- Media : Dry Block
- Operating Temp : 0 - 30 ° C

ELECTRICAL SPECIFICATION

- Power consumption : 1000 Watts approx.
- Power supply : 230 Vac, 50 Hz
- Fuse rating : 10A, 250 Vac
- Cooling system : Air cooling through fans

PHYSICAL SPECIFICATION

- Dimensions : 5.5" (W) X 14" (H) X 12" (L)
- Weight : 7.5 Kg. approx.
- Well adapter M. O. C. : Aluminium
- Well Diameter : Max. up to 16 mm
- Well Depth : 150 mm
- Accessories : Set of insert / pullout tool,
(7mm/8.5mm/10mm/12.5mm/16mm), Carrying Bag

MILLENNIUM INSTRUMENTS LIMITED

H.O. : 101, 'Dwarkashish', Main Vikas Marg, Shakarpur, Delhi-92 (INDIA)
 Tel. : +91-11-2201-4325, 4327, 65094516. Fax.: 22014326
 Website : www.mil.org.in , www.miepl.com, E-mail : miepl@miepl.com

Works : F-12, Site-B, UPSIDC Ind. Area Surajpur, Greater Noida,
 Gautam Budh Nagar-201306 (U.P.) E-mail : sales@miepl.com, info@miepl.com